



BIDS AND AWARDS COMMITTEE I

Supplemental/Bid Bulletin No. 1

SUPPLY AND DELIVERY OF ONE (1) UNIT BRAND NEW ENT OPERATING MICROSCOPE FOR THE PHILIPPINE ARMY (PA)

Bid Reference No. MPG-B1-2019-161
Approved Budget for the Contract – ₱3,678,846.15

This **Supplemental/Bid Bulletin No. 1** is being issued to amend some provisions of the Bidding Documents and response to queries raised by suppliers during the Pre-Bid Conference on 18 July 2019 and through letters/email for the aforesaid project.

A) NEW SCHEDULE FOR THE SUBMISSION AND OPENING OF BIDS

27 August 2019; 2:30PM

B) REPLY TO QUERIES

Queries from MICROLAB PHILIPPINES		
PA Technical Specification	Microlab Philippines clarifications on the Technical Specification	Response
Objective Lenses: a. 200mm b. 300mm c. 400mm	a. 200mm b. 250mm c. 400mm Explanation: We are offering 250mm because this is intended for ENT surgery. The 300mm is a working distance used for Neuro Surgery cases.	Request denied. The Philippine Army TWG for Medical Items retains the original requirements.
Binoculars: Straight inclinable tiltable by 60-80degrees Viewing tube: F= 160 – 180 mm	Straight inclinable tiltable by 60-80 degrees or better Viewing tube: F = 160 – 180 mm Explanation: We can offer a binocular that can be inclined and tilted up to 180 degrees	Request denied. The Philippine Army TWG for Medical Items retains the original requirements.

Eyepiece: a. Widefield b. High eye point c. Integrated eyecups d. Minimum magnification – 12.5x e. Manually adjustable interpupillary distance	a. Widefield b. High eye point c. Integrated eyecups d. Minimum magnification – 10x or 12.5x e. Manually adjustable interpupillary distance	Request denied. The Philippine Army TWG for Medical Items retains the original requirements.
Magnification changer: a. Five (5) Step manually adjustable b. Magnification factor – 0.4, 0.6, 1.0, 1.6 & 2.4	a. Five (5) Step manually adjustable b. Magnification range: 2.3x to 22.2x	Request denied. The Philippine Army TWG for Medical Items retains the original requirements.
Focusing: a. Manual fine focusing b. Micro-range of 40-45cm c. Foot pedal controlled focus	a. Manual focusing b. Micro-range of 40-45cm c. Foot pedal controlled focus	Request denied. The Philippine Army TWG for Medical Items retains the original requirements.

This Supplemental/Bid Bulletin 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 14th day of August 2019 in Makati City.

Reviewed and Approved by:

CHRISTABELLE P. EBRIEGA
Chairperson, Bids and Awards Committee – I

ATTY. MARIA GUDELIA C. GUESE
Vice Chairperson

MYRA CHITELLA T. ALVAREZ
Member

DAVID A. INOCENCIO
Member

JOEL S. RODRIGUEZ
Member

Concurred by:

BGEN GLENN CRUZ AFP
Alternate Provisional Member- PA

Received by:	
_____ (SIGNATURE OVER PRINTED NAME & DATE)	_____ NAME OF COMPANY
(PLEASE RETURN OR FAX THIS PAGE ONLY TO THE PITC BAC-I)	