



BIDS AND AWARDS COMMITTEE I

Supplemental/Bid Bulletin No. 1

SUPPLY AND DELIVERY OF 119 UNITS BRAND NEW HOSPITAL BED FOR THE PHILIPPINE ARMY (PA)

Bid Reference No. MPG-BI-2020-235

Approved Budget for the Contract – ₱5,721,153.48

This **Supplemental/Bid Bulletin No. 1** is being issued to further clarify, modify and amend the provisions in the Bidding Documents and in response to clarifications from prospective bidder/s for the aforementioned project.

REPLY TO QUERIES OF BIDDERS:

Below are the queries/requests for consideration/clarification by the prospective bidders, which have been answered by the PITC Bids and Awards Committee I and Philippine Army (PA).

Query from BEROVAN MARKETING, INC. per letter emailed dated 28 September 2020			
Query 1:	Parameter	Description	Clarification
	Frame	1. Steel Epoxy 2. Polyester powder coated	1. Perforated Epoxy 2. Epoxy Powder Coated
PA's reply	<i>Request Denied.</i> The TWG-Medical Items retain the original requirements.		
Query 2:	Parameter	Description	Clarification
	Mattress	Rubberized with leatherette Cover	Foam with leatherette Cover
PA's reply	<i>Request Denied.</i> The TWG-Medical Items retain the original requirements.		
Query 3:	Parameter	Description	Clarification
	Functions	1. Height (adjustable) 420-720mm 2. Back-Rest lifting angle ± 5 0-75° 3. Knee-rest lifting angle ± 5 0-45°	1. Height (adjustable) – without mattress 500-750mm 2. Back-Rest lifting angle 80° ± 5 3. Knee-rest lifting angle 45° ± 5
PA's reply	The suggested range is within the original requirements.		
Query 4:	Parameter	Description	Clarification
	Crank handles	Steel Chrome	ABS plastic
PA's reply	<i>Request Denied.</i> The TWG-Medical Items retain the original requirements.		



Query 5:	Parameter	Description	Clarification
	Accessories	1. One (1) piece Dining Table 2. Four (4) pieces drip holes	1. One (1) piece Overbed table with wheels 2. Is this the hole for IV hole? 1 hole in each corner of the bed?
PA's reply	Request Denied. 1. The TWG-Medical Items retain the original requirements. 2. Yes, it is intended for the IV Pole.		
Query 6:	Parameter	Description	Clarification
	Dimensions: LxWxH	2,150-2,300mm x 950-1,000mm x 720-750mm	2,080 x 970 x 500mm
PA's reply	Request Denied. The TWG-Medical Items retain the original requirements.		

Query from R&A DENTAL SUPPLY CENTER CO. per letter emailed dated 08 October 2020	
Query 1:	May I be clarified is what type of dining table to be complied for the Accessories: Item No. 1. One (1) piece Dining Table.
PA's reply	Standard/generic hospital dining table.
Query 2:	May I be clarified is what does it mean for Mattress: Rubberized with Leatherette Cover.
PA's reply	A Rubberized mattress is firmer than a foam mattress and the leatherette cover for the mattress is a plastic-based vinyl material (synthetic type) that would serve to repel food and liquids, scratches and can be washable.

This **Supplemental/Bid Bulletin No. 1** shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 12th day of October 2020 in Makati City.

Reviewed and Approved by:

CHRISTABELLE P. EBRIEGA
Chairperson, Bids and Awards Committee – I

MYRA CHITELLA T. ALVAREZ
Vice Chairperson

JOEL S. RODRIGUEZ
Member

ATTY. MITZELL ARTHUR R. MAGDAONG
Member



ATTY. ROXANNE MARIE Q. CRUZ
Member

ATTY. MARK BRIAN A. DELA CRUZ
Alternate Member

Concurred by:

MGEN GLENN E CRUZ PA
Provisional Member- PA

Received by:	
<hr/>	<hr/>
(SIGNATURE OVER PRINTED NAME & DATE)	(NAME OF COMPANY)
(PLEASE RETURN OR ELECTRONIC MAIL THS PAGE ONLY TO THE PITC BAC)	